



EC Certificate
Directive 93/42/EEC Annex V
Production Quality Assurance
Medical Devices

Registration No.: DD 60140740 0001

Report No.: 15070840 008

Manufacturer: V&Q Manufacturing Corporation
4A Fuyuan Road, Pengchang Town
Xiantao City
433018 Hubei
China

Products: Medical Devices
(see attachment for details)

Replaces: [blank] No.: DD 60122784 0001

Expiry Date:

The Notified Body has verified that the requirements of Annex V of the directive 93/42/EEC have been met for the named manufacturer has established and applies a quality assurance system and periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. The market of class IIb and class III devices covered by this certificate according to Annex III is required.

Effective from: 2019-07-04

2019-07-04



TÜVRheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

TÜVRheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

Flammability of Clothing Textiles Final Report

Test Article: VQN0185W-3
 Study Number: 947950-S01
 Study Received Date: 22 Feb 2017
 Test Procedure(s): Standard Test Protocol (STP) Number: 801-STP0073 Rev 06

Summary: This procedure was performed to evaluate the flammability of plain surface textile fabric by measuring the ease of ignition and the speed of flame spread. The purpose of this test is to separate materials into different classes, thereby assisting in a judgment of the suitability of the material and protective clothing material. The test procedure was performed in accordance with the test method outlined in 16 CFR Part 1610 (a) *Step 1 - testing in the original state*. The *Step 2 - testing after refurbishing*, was not performed. All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practices for medical devices under CFR Parts 210, 211 and 820.

Test Article Side Tested: Outside Surface
 Orientation: Machine

Test Criteria for Specimen Classification (See 16 CFR 1610.10)

Class	Criteria
1	Flame spread not observed on plain surface Textile Fabric
2	Burn time ≥ 3.5 seconds
3	Flame spread not observed on plain surface textile fabrics
3	Burn time < 3.5 seconds

The 16 CFR Part 1610 standard requires that test articles are to be tested if, during preliminary testing, only 1 test article exhibits flame spread greater than 3.5 seconds or the test articles exhibit an average flame spread less than 3.5 seconds. For each test article, 5 replicates are to be tested if no flame spread is observed upon preliminary testing. If a test article exhibits flame spread and it is equal to or greater than 3.5 seconds, or the average flame spread is equal to or greater than 3.5 seconds. In accordance with the standard, 5 replicates are to be tested for this study.

Results:

Replicate	Time of Flame Spread
1	DNI
2	DNI
3	DNI
4	DNI
5	DNI

DNI = Test article did not ignite

Janelle Benty for
 Study Director
 Brandon L. Williams

07 Mar 2017
 Study Completion Date



947950-S01

Synthetic Blood Penetration Resistance Final Report

Test Article: VQN0185W-3
Study Number: 947953-S01
Study Received Date: 22 Feb 2017
Test Procedure(s): Standard Test Protocol (STP) Number: 801-STP0012 Rev 06

Summary: This procedure was performed to evaluate surgical facemasks and clothing materials designed to protect against fluid penetration. The purpose was to simulate an arterial spray and evaluate the effectiveness of the test article from possible exposure to blood and other body fluids. The distance from the tip of the cannula is 30.5 cm. A test volume of 2 mL of synthetic blood was used using plate method.

This test method was designed to comply with ASTM F2100 (as referenced in EN 14683:2014) with the following exception. ISO 22623 was performed in an environment with a temperature of $21 \pm 5^\circ\text{C}$ and a relative humidity of $85 \pm 5\%$. Instead, testing was performed at ambient conditions within one minute of the environmental chamber held at those parameters.

All test method acceptance criteria were met in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR 312.620.

Number of Test Articles Tested: 2
Number of Test Articles Passed: 2
Test Side: A
Pre-Condition: at $21 \pm 5^\circ\text{C}$ and $85 \pm 5\%$ relative humidity (RH)
Test Condition: RH

Results: Per ASTM F2100, an acceptable quality limit of 4.0% is met for a normal single sampling plan which shows passing results.

Test Article	Pressure (mm Hg)	Result
VQN0185W-3	80	None Seen


Study Director Brandon L. Williams


Study Completion Date



Latex Particle Challenge Final Report

Test Article: VQN0185W-3-1
 VQN0185W-3-2
 VQN0185W-3-3
 VQN0185W-3-4
 VQN0185W-3-5
 Study Number: 947951-S01
 Study Received Date: 22 Feb 2017
 Test Procedure(s): Standard Test Protocol (STP) Number: 801-STP0005

Summary: This procedure was performed to evaluate the non-viable particle count (NVC) of the test article. Monodispersed polystyrene latex spheres (PSL) were used as challenge particles and passed through the test article. The particles that passed through the test article were counted using a laser particle counter.

Three one-minute counts were performed, with the test article in place. The results were averaged. Three one-minute control counts were performed, without the test article in place, before and after each test article and the counts were averaged. Control counts were used to determine the average number of particles delivered to the test article. The filtration efficiency was calculated using the average number of particles penetrating the test article compared to the control values.


The procedure employed the basic particle filtration test method in ASTM F2299, with some exceptions; notably the procedure incorporated a neutralized aerosol. In real use, particles carry a charge, thus this challenge represents a more realistic challenge. A neutralized aerosol is also specified in the FDA guidance document on surface disinfection. Method acceptance criteria were met. Testing was performed in compliance with Good Manufacturing Practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Site: [Redacted]
 Area: [Redacted]
 Parameters: relative humidity (RH) at 0758; 21°C, 23% RH at 1030
 Laboratory: [Redacted]
 Average Filtration Efficiency: [Redacted]

Results:

Test Article	Test Article Counts	Average Control Counts	Filtration Efficiency (%)
VQN0185W-3-1	35	13,428	99.74
VQN0185W-3-2	34	13,140	99.74
VQN0185W-3-3	31	13,057	99.76
VQN0185W-3-4	31	13,190	99.76
VQN0185W-3-5	42	13,286	99.68


 Study Director Brandon L. Williams


 10 Mar 2017
 Study Completion Date



Certificate

The Certification Body of
TÜV Rheinland LGA Products GmbH

hereby certifies that the organization
V&Q Manufacturing Corporation
4A Fuyuan Road, Pengchang Town
Xiantao City
433018 Hubei
China

has established and applies a quality management system for the following products: medical devices

Manufacture and Distribution of Medical Devices
(see attached list of products included)

Proof has been provided that the organization meets the requirements specified in

ISO 13485:2016

are subject to yearly surveillance. The management system is subject to yearly surveillance.

Effective date: 2018-02-12
Certificate No.: SX 60122785 0001
Product group defined. Report No.: 15070840 005
Certificate is valid until: 2020-07-24

Certification Body



Date 2018-02-12

AAB

PROTECTIVE BREATHING

MASK

Anti-fog/Anti-droplets/Anti-dust



Long-lasting
Effective
Breathable

10 pcs

KN95 ffp2 EN 149:2001



THE KEITH & EVEN GROUP™
SAMPLE SAMPLE
THE KEITH & EVEN GROUP™

AAB

Method of use



1. Keep your hands dry and unfold the mask.
2. Keep the white side to your face and the nose clip along to the upside.
3. Hang on to your face with the ear loops and adjust the both loops even, then press the nose clip.
4. Adjust the facemask to the most fit of your face in each side.

- Notes
- Wash your hands before wearing the mask and avoid any touch to your face during wearing.
 - Keep the mask flat and the inner surface to the upside after taking off, and fold it slightly and keep in the shade when necessary to avoid the breaking the nose clip.
 - Do not use when you have dizziness or too deep.
 - Not recommended to the children under 14 years old.

AAB

PROTECTIVE BREATHING

MASK

Anti-fog/Anti-droplets/Anti-dust





CE Documentation Review

No. 38200391T.HKM0503

Holder: Hunan Kanglilai Medical Instrument Co., Ltd.

108 Longyang South Road, Longyang Community, Gangliang Street, Hangzhou City, Hunan Province, China

Review goal: Verification of the Technical Documentation of Technical Devices Drawings with the Medical Devices Directive Annex VII

Product: Medical Device (Non sterile)

Model(s):

Classification: Class II (sterile)

According to the Manufacturer's Declaration

Review: This document has been issued on a voluntary basis and upon request of the manufacturer, it is our opinion that the Technical Documentation shared with us by the manufacturer is compatible with the European Standard for Medical Devices. Technical documentation identified with the no. CE-FM-(01-15).

The manufacturer is responsible for the CE Marking process, and not exempted to carry out all necessary compliance activities. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01_ECM rev.3 available at: www.entecema.it

Date of issue 31 March 2020

Approver
ECM Service Director
Luca Pedroni



Expiry date 30 March 2025

Technical Expert
Amanda Piro



検査報告-검토 보고서- Rapport d'Evaluation

Review Report



Verification of Conformity

Applicant: V&Q Manufacturing Corporation
Address: #B1614 Optical Valley Times Square Guanshandada
Wuhan City, Hubei Province China
Product(s): Surgical mask
Type(s): VQN0185B-3, VQN0186G-3

The submitted sample of the above product has been tested in accordance with the related European standards EN 14683:2014 Surgical Masks and related methods.

Test Information:
Laboratory: Nelson Laboratories, Inc.
Test Report(s) No.: 957829-S01, 947953-S01
Type: Bacterial Filtration Efficiency (BFE) and Differential Pressure (Delta P) Medical Masks
Result: Synthetic Particle Filtration Efficiency (SPE) Type IIR

The test result and technical specifications of the listed devices is conformity with the requirements of Medical Devices and related standard.

SUNGO Certification Limited

THE KEITH& EVEN GROUP™ THE KEITH& EVEN GROUP™ THE KEITH& EVEN GROUP™
SAMPLE SAMPLE SAMPLE

